



Vitrek Corporation
 12169 Kirkham Road
 Poway, CA 92064
 Voice: 858 689 2755 • Fax: 858 689 2760 • www.vitrek.com

Report #:
 Model #:
 Serial #:

Outgoing Measurement Results

DIRECT:COMMON CALIBRATION VERIFICATION RESULTS							
NOM. INPUT (V)	FREQUENCY	ACTUAL INPUT (V)	READING (V)	ERROR (V)	LIMITS (V)	UNCERTAINTY (ppm)	STATUS
0	DC		0.01	0.01	0.04	N/A	PASS
0	AC		0.02	0.02	0.11	N/A	PASS
100	DC	100.0	100.01	0.01	0.07	15	PASS
200	DC	200.0	200.01	0.01	0.10	27	PASS
1000	DC	1000.0	1000.03	0.03	0.34	26	PASS
-1000	DC	-1000.0	-999.99	0.01	0.34	26	PASS
1000	50/60Hz	1000.0	999.97	0.03	1.31	43	PASS
1000	400Hz	1000.0	998.0	2.0	4.2	43	PASS
4000	DC	4001.8	4001.9	0.1	1.4	73	PASS
9000	DC	8992.1	8992.4	0.3	3.9	73	PASS
9000	50/60Hz	9000.0	9002.3	2.3	12.1	280	PASS

DIRECT:COMMON COMMON MODE VERIFICATION RESULTS							
NOM. INPUT (V)	FREQUENCY	ACTUAL INPUT (V)	READING (V)	ERROR (V)	LIMITS (V)	UNCERTAINTY (ppm)	STATUS
500	DC		0.04	0.04	0.29	N/A	PASS
500	50/60Hz		0.81	0.81	2.76	N/A	PASS

HI PROBE INPUT STANDARDIZATION CALIBRATION							
NOM. INPUT (μA)	FREQUENCY	ACTUAL INPUT (μA)	READING (μA)	ERROR (μA)	LIMITS (μA)	UNCERTAINTY (ppm)	STATUS
0	DC		0.0001	0.0001	0.0003	N/A	PASS
0	AC		0.0002	0.0002	0.0007	N/A	PASS
1	DC	0.9950	0.9950	0.0000	0.0005	98	PASS
2	DC	1.9900	1.9901	0.0001	0.0006	98	PASS
10	DC	9.950	9.950	0.000	0.002	98	PASS
20	DC	19.900	19.900	0.000	0.003	98	PASS
100	DC	99.50	99.50	0.00	0.02	98	PASS
200	DC	199.00	199.00	0.00	0.03	98	PASS

LO PROBE INPUT STANDARDIZATION CALIBRATION							
NOM. INPUT (μA)	FREQUENCY	ACTUAL INPUT (μA)	READING (μA)	ERROR (μA)	LIMITS (μA)	UNCERTAINTY (ppm)	STATUS
0	DC		0.0000	0.0000	0.0003	N/A	PASS
0	AC		0.0002	0.0002	0.0007	N/A	PASS
1	DC	0.9950	0.9949	0.0001	0.0005	98	PASS
2	DC	1.9900	1.9898	0.0002	0.0006	98	PASS
10	DC	9.950	9.950	0.000	0.002	98	PASS
20	DC	19.900	19.900	0.000	0.003	98	PASS
100	DC	99.50	99.50	0.00	0.02	98	PASS
200	DC	199.00	199.00	0.00	0.03	98	PASS

* : Indicates possibility of falling within or outside of the specified limits due to the measurement uncertainty.